

Device ID: 21620

Eng ID: 105.1N.0014

Scanned:

**NEPTUNE Canada (MTC) Instrument Functional Test Rev B**

3.3.1 Identification

Manufacturer: GURALP SYSTEMS

Model Number: CMG NEPTUNE OBS Serial Number: 2216

3.3.2 Status

Active  Receiving - passed  Incoming Inspection - passed

3.3.3 Vendor Acceptance Test

Contact Name: Number: Called:

Question:

Immersion Required? Yes  No

3.3.4 Initial Test Set-up

Power and Comms OK

PERFORMED BY BOB MELDRUM.

Sensors/ Options:

Data Format OK  Problem:

3.3.5 Inrush Test (type test)

SEE S/N 2217

Test Voltage: 48.0 Quiescent Current: 40mA

Waveforms saved  Peak Inrush: 15A

3.3.6 Ethernet Comms Test

Ethernet device - IP address: Subnet mask:  
Gateway: Comms  Remote Config

3.3.7 Isolation Test (Wet Test)

NOT PERFORMED DUE TO TIME CONSTRAINTS.

Operation while immersed  insulation resistance value:

3.3.8 Final Test

NOT PERFORMED DUE TO TIME CONSTRAINTS

JB#: JB port#: Cable ass'y:

Power and Comms OK

'Ground Fault' value: Current:

2009.09.10 (yyyy.mm.dd)

Completion noted in DMAS

date

signature R. Riddell

Technician: RICHARD RIDDELL

signature: R. Riddell