

Device ID: 22057 Eng ID: 103.1N.0007 Scanned:

NEPTUNE Canada (MTC) Instrument Functional Test Rev A

3.3.1 Identification

Manufacturer: GURALP

Model Number: CMG NEPTUNE OBS Serial Number: 2217

3.3.2 Status

Active Receiving - passed Incoming Inspection - passed

3.3.3 Vendor Acceptance Test

Contact Name: Question: Immersion Required? Yes No

3.3.4 Initial Test Set-up

Power and Comms OK

PERFORMED BY BOB MELDRUM

Sensors/ Options:

Data Format OK Problem:

3.3.5 Inrush Test (type test)

Test Voltage: 48.0 Quiescent Current: 40 mA

Waveforms saved Peak Inrush: 15 A

3.3.6 Ethernet Comms Test

Ethernet device - IP address: Subnet mask: Gateway: Comms Remote Config

3.3.7 Isolation Test

NOT PERFORMED DUE TO TIME CONSTRAINTS.

Operation while immersed insulation resistance value:

3.3.8 Final Test

NOT PERFORMED DUE TO TIME CONSTRAINTS.

JB port#: Cable ass'y: Power and Comms OK

'Ground Fault' value: Current:

2009.09.10 (yyyy.mm.dd) date

Completion noted in DMAS

signature R. Riddell

Signature R. Riddell