

Instrument Reception And Incoming Inspection Checklist

Rev A - Pre-release

ESD control measures exception:

Identification and Shipping Documentation 3.3.1

Manufacturer: *ASL Environmental Sciences Inc*Vendor (If Different): *~~ASL~~*Model Number: *AWCP5 - WCP16*Serial Number: *5016*Principal Investigator: *Dower*Purchasing Agent (If Different): *UUC*Target Deployment Site: *Barkley Canyon UPS*

Shipping Damage 3.3.2

Describe any evidence of damage or tampering in shipping or handling. Append digital images where appropriate. *

well packed & intact

Completeness of Shipment 3.3.3

Note any missing material or other deviation from the packing list. *

*Packing list missing - Drives (spares),
USB-serial adapter, Quick Start sheets missing,
Vendor acknowledged and will supply - accepted*

Conformance with Requirements 3.3.4

Indicate compliance with the following requirements:

Package marked ESD sensitive?

Yes No

Purpose-built, reusable shipping enclosure?

Yes No Fail

Instrument fixed in position (no loose packing)?

Yes No Fail

Package marked with shock constraint?

Yes No

Shock recording device enclosed?

Yes No Fail

Allowable shock limit exceeded? *

Yes No Fail

Connectors covered and protected?

Yes No Fail