SAT-QR-95001 Verification Checklist In Situ Ultraviolet Spectrophotometer (ISUS)

Quality Assurance:

SATLANTIC INCORPORATED



This document is the final inspection, verification checklist for a Satlantic ISUS. The inspection is performed after all other work has been completed, before it is shipped to a customer. The inspection is conducted and QA counter signed by qualified Satlantic Technicians.

Serial Number: Version: Construction: Battery Pack: Project Number: Date:	Standard Aluminum SN: SN: 2009-277 1) Aug. 20	ISUS-X PVC
Interface Functions (ISUSCom): Ay Check instrument setup		
A Confirm default schedule present		
Run complete system (cables, battery packs, MDU)		
Run acquisition		
イタ Create Summary Report		
Store internal copy		
Test update calibration procedure		
Data Functions:		
Nitrate value using DIW		0.49 µM
Nitrate value using ASW		-0.44 µM
Nitrate value using standard solution		80.31 µM
Instrument temperature after 10 minutes		28.5 °C
Lamp temperature after 10 minutes		41.2 °C
Spec temperature after 10 minutes		<u>24,1 °C</u>
Internal humidity value		6.5 %
Shutter function - check dark frames		860 Counts
Test binary output		
Delete test files from disk		
Handrian Francisco		
Hardware Functions: Test analog output. Voltage at midscale		2.061 v
Test scheduled mode		
Reset to continuous		
Run ISUS from 2-pin Power port. Current Draw at 12V		540 mA
Run ISUS from 6-pin P/T port. Current Draw at 25V		300 mA
Check cable terminations, splices, and pigtails		

+1.902.492.4780

Date: 12 Aug 2010

satlantic.com

info@satlantic.com