


PO Box 518  
620 Applegate St.  
Philomath OR 97370  
wetlabs@wetlabs.com



(541) 929-5650  
Fax (541) 929-5277  
<http://www.wetlabs.com>

Date: 10/24/13

Customer: University of Victoria

S/N#: WQM-233

CTD PSI: 100m

ECO Range: CHL: 50 NTU: 25

**Repairs and Modifications:**

Replaced the CTD Assembly.

WETLabs serviced the ECO Meter.

Replaced the BLIS Assembly.

Replaced the 9V UPS Battery and Coin Cell.